

Northbrook, Illinois • (847) 272-8800
 Melville, New York • (631) 271-6200
 Santa Clara, California • (408) 985-2400
 Research Triangle Park,
 North Carolina • (919) 549-1400
 Camas, Washington • (360) 817-5500



MASCOT ELECTRONIC A/S
 UL COORDINATOR-ENGINEERING SEC
 MOSSEVEIEN 109
 PO BOX 177
 1601 FREDRIKSTAD NORWAY

RE: Project Number(s) - 00SC03378

Your most recent listing is shown below. Please review this information and report any inaccuracies to the UL Engineering staff member who handled your project.

For information on placing an order for UL Listing Cards in a 3 x 5 inch format, please refer to the enclosed ordering information.

QQHM2 March 16, 2001
Power Supplies, Medical And Dental - Component

MASCOT ELECTRONIC A/S
MOSSEVEIEN 109 PO BOX 177, 1601 FREDRIKSTAD NORWAY

E208102

Model No.	Rated Input		SC	Output			OC	SP	EP	FC	GC
	V	Hz		Max V	Max A	Max VA					
9920	100-250	50-60	0	12	3	40	9	2601	15B	6,7	1
				16	2.5	40					
				24	1.6	40					
				48	0.8	40					
9920, Ta>30C				12	2.5	32					
				16	2	32					
				24	1.3	32					
				48	0.6	32					
9921	100-250	50-60	0	12	3	40	9	2601	15B	6,7	1
				16	2.5	40					
				24	1.6	40					
				48	0.8	40					
9921, Ta>30C				12	2.5	32					
				16	2	32					
				24	1.3	32					
				48	0.6	32					
9922	100-250	50-60	0	12	3	40	9	2601	15B	6,7	1
				16	2.5	40					
				24	1.6	40					
				48	0.8	40					
9922, Ta>30C				12	2.5	32					
				16	2	32					
				24	1.3	32					
				48	0.6	32					

Marking: Company name and model designation.

See General Information Preceding These Recognitions